



"हरित नगर, समृद्ध दमक"

दमक नगरपालिका

नगर कार्यपालिकाको कार्यालय



प.सं. ०६९१०८०

दमक, झापा

०२३-२६००४४७६०१३०७६०४०२७६१०७३


च.न ८५४५

कोशी प्रदेश, नेपाल

फ्याक्स नं. ०२३-२६०१४६

www.damakmun.gov.np

info@damakmun.gov.np



मिति : २०८०।०३।७

श्री सम्पूर्ण सुचिकृत फर्महरु

विषय : दर रेट पेश गर्ने सम्बन्धमा ।

उपरोक्त सम्बन्धमा यस दमक नगरपालिकाका नागरिकहरुलाई सुस्तश्रवण शक्ति यन्त्र वितरण गर्नु पर्ने भएकाले यसै पत्रसाथ उल्लेखित स्पेसिफिकेशन अनुसारको सामग्रीहरुको दररेट मिति २०८०।३।२९ गते १२ वजे भित्र दमक नगरपालिकामा पेश गर्नुहुन अनुरोध छ । पेश हुन आएका दररेट सोहि दिन १ वजे खोलिने जानकारीको लागि समेत अनुरोध छ ।

आवश्यक कागजातहरु

2080/3/6
उमरु प्रसाद निरौला
मुख्य प्रशासकीय अधिकृत

१. फर्म दर्ता प्रमाण पत्रको प्रतिलिपी

२. मुअकर दर्ता प्रमाण पत्रको प्रतिलिपी

३. अधिल्लो आ.व.को कर चुक्ता

४. दररेट समावेश भएको कागजात

५. सामग्रीको नमुना (Brand, Model no, specification) आदी सम्पूर्ण उल्लेख हुनु पर्ने छ ।

६. १ वर्षको वारेन्टी declaration

७. Manufacture Authorization

TECHNICAL SPECIFICATION FOR HEARING AID

S.N.	Purchaser's Specifications	Bidder's compliance sheet		
		yes/No	Page no. in catlog	Remarks
	HEARING AID			
	Manufacturer			
	Brand			
	Type / Model			
	Country of Origin			
1	Description of Function			
	Ear Machine for Hearing Aid provides the required aid when all hear is void.It is made of medical-grade material that allows to safely use it for more than 6-8 hours			
2	Operational Requirements			
	simple device designed to hide right behind your ear, actually a major breakthrough in sound interception and amplification technology.			
3	System Configuration			
	Hearing aid with complete unit			
4	Technical Specifications			
A	Ensures optimum amplification and hearing			
B	Ergonomic design for great comfort			
C	70dB gain, 13 size battery or equivalent, 6 signal processing channel, 4 hearing profile, Maximum output 132dB			
D	anti feedback system adaptive			
E	noise reduction			
F	Volume can be adjusted.			
G	Replacable batteries			
5	Standards and Safety Requirements			
A	Must submit ISO for medical device and CE or USFDA approved			
6	User Training			
	The Supplier shall conduct user/technician training for this equipment to enable operators to use the equipment properly.The training shall include the use of all operational functions of the equipment,as well as routine checks and maintenance and maintenance expected by users			
7	Warranty			
	Comprehensive warranty for 1 year after acceptance			
6	Operating Environment			
A	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country.			
9	Documentation and accessories			
	Should provide User Manual and one piece of complete hearing aid during technical evaluation			

Bidder must completely fill the Technical Specification Form (TSF). Only Yes/no/all complies should not be written. Pagenumber in the catalogue of all the required parameters must be clearly mentioned and highlighted. Failure in doing so may lead to rejection of bid from technical committee

Naependra Oli

[Signature]

नगर कार्यपालिका
 वसन्त-३
 कोशी प्रदेश
 २०७३

गौला